

ABSTRACT

Dose distribution is measured in a patient-shaped phantom with high accuracy for radiation therapy treatment verification. A detector configuration in such a phantom and improvements in measurement methodology enable application of correction factors in an accurate way.